

URGENT MEDICAL DEVICE CORRECTION



3 June 2026

GE HealthCare Ref. # 32101

To: Healthcare Administrator / Risk Manager
Director of Neonatology/L and D/Nurse Manager
Director of Biomedical Engineering

RE: **Canopy Soffits and fasteners on Giraffe OmniBed and Giraffe OmniBed Carestation Devices**

Safety Issue

GE HealthCare has become aware of a potential issue involving the canopy soffit beneath the control panel end of the device. Under certain conditions, for example, impact during transport, one or more of the fasteners that secure the soffit can disengage. If this occurs, a fastener could fall into the patient compartment. In rare circumstances, if multiple fasteners disengage, the canopy soffit can become loose or detach. These conditions could potentially result in patient injury.

There have been no injuries reported to GE HealthCare as a result of this issue.

Actions to be taken by Customer/User

You may continue to use your device after performing the pre-use checkout in accordance with the Giraffe OmniBed Carestation User Manual (Chapter 6 – Pre-Use Checkout).

During the pre-use checkout described in the product User Manual, visually inspect the device for any damaged or missing components. This includes verifying that the canopy soffit is properly secured and that all six fasteners are present and fully engaged. See Figure 1.




Figure 1: OmniBed Canopy Assembly

If you observe damage, missing fasteners, or improper soffit attachment during the pre-use checkout, please remove the device from clinical use and contact GE HealthCare Service.

Please ensure all potential users in your facility are made aware of this safety notification and the recommended actions.

Please complete and return the attached acknowledgement form electronically via [FMI 32101 Digital CRF](#) or print, fill out manually, scan, and email to MIC.FMI32101@gehealthcare.com.

Please retain this document for your records.

Affected Product Details

Product	Model #	GTIN Number
Giraffe OmniBeds*	All Models	*See Note Below
Giraffe OmniBed Carestations	2082844-001-XXX	010084068211686221
Field Replaceable Unit (FRU)	6600-1056-400	NA
Field Replaceable Unit (FRU)	6600-1461-500	NA

*NOTE: Some products were shipped prior to the implementation of UDI and may not contain a Global Trade Item Number (GTIN)

Intended Use:

The Giraffe OmniBed and Giraffe OmniBed Carestation are a combination of an infant incubator and an infant warmer. The device can be operated as an incubator or as a warmer and can transition from one mode to the other on user's demand. It cannot be operated in both modes at the same time. Incubators and warmers provide heat in a controlled manner to neonates who are unable to thermo-regulate based on their own physiology. Incubators provide an enclosed, temperature-controlled environment and warmers provide infrared heat in an open environment. They may also be used for short periods of time to facilitate the neonate's transition from the uterus to the external environment. This device may incorporate a Servo Controlled Oxygen Delivery System. This is indicated to provide stable oxygen concentration within the infant compartment at the value set by the operator (21-65%).

Product Correction

GE HealthCare will correct all affected products at no cost to you. A GE HealthCare representative will contact you to arrange for the correction.

Contact Information

If you have any questions or concerns regarding this notification, please contact GE HealthCare Service or your local Service Representative.

Please be assured that maintaining a high level of safety and quality is our highest priority. If you have any questions, please contact GE HealthCare per the contact information above.

Sincerely,

A handwritten signature in blue ink, appearing to read 'Laila', followed by a long horizontal flourish.

Laila Gurney
Chief Quality & Regulatory Officer
GE HealthCare

A handwritten signature in blue ink, appearing to read 'Sk', followed by a horizontal flourish.

Scott Kelley
Chief Medical & Safety Officer
GE HealthCare

MEDICAL DEVICE NOTIFICATION ACKNOWLEDGEMENT
RESPONSE REQUIRED

Please complete this form and return it to GE HealthCare promptly upon receipt and no later than 30 days from receipt. This will confirm receipt and understanding of the Medical Device Correction Notice.

Facility Name: _____

Street Address: _____

City/State/ZIP/Country: _____

Customer Email Address: _____

Customer Phone Number: _____

By signing this form, we acknowledge receipt and understanding of the accompanying Medical Device Notification, and that we have informed all potential users and have taken and will take appropriate actions in accordance with that Notification.

Please provide the name of the individual with responsibility who completed this form.

Signature: _____

Printed Name: _____

Position/Job Title: _____

Date (DD/MM/YYYY): _____

To complete this form electronically, please scan the QR Code below or click this link:

<https://gehealthcare-svc.my.site.com/publicForm/s/?formId=aGjUr00000383yH>



To complete this form via email, scan or take a photo of the completed form and email to:

MIC.FMI32101@gehealthcare.com

